### TSC TLIF Banana

Diffe

-Da

Surgical Technique Manual




## VIEWS & SIZES



To accommodate the various anatomical ranges, the TSC TLIF Banana cage is available in a wide variety of heights and endplate footprints

#### **MEDIUM / LARGE**

		Height (mm)								
		7	8	9	10	11	12	13	14	15
Length / Lordosis	28 mm/ 6°	✓	~	✓	✓	✓	✓	✓	✓	✓
	32mm / 6°	✓	~	$\checkmark$	✓	✓	✓	✓	✓	~
	28 mm/ 12°	✓	~	✓	✓	✓	✓	✓	✓	✓
	32mm / 12°	✓	✓	✓	✓	✓	✓	$\checkmark$	✓	✓
	28 mm/ 18°	×	✓	✓	✓	✓	✓	✓	✓	✓
	32mm / 18°	×	>	~	$\checkmark$	$\checkmark$	$\checkmark$	✓	$\checkmark$	>



## INSTRUMENTS





Impactor

Spoon Curette

#### Ring Curette



 $\sim$ 



Shavers

T-Handle

Retractor

Hammer

#### Introduction

#### Material:

The TSC PLIF / TLIF cages are made of PEEK by Invibio (Poly Ether Ether Ketone ) is a radiolucent material with an elastic modulus close to bone. The TSC Cage family implants are covered with Orthobion's unique thin film titanium coating.

#### **Characteristics:**

The design and form of the TSC Cages is close to the shape of the distal disc space shape.

This design leads to :

- primary fixation due to a specific parabolic profile cranial and caudal of the endplates of the cage
- preventing possible pull our movements
- thin titanium film coating for better promotion of osseous regrowth and on growth to favorise a stable solid arthrodesis.
- preventing of withdrawal movements
- correction of intervertebral disc height
- restoring physiological lordosis

#### Indications:

- Degenerative Disc Disease (DDD) at levels from S1 till L2
- DDD defined by:
- Discogenic back pain
- Grade 1 Spondylolisthesis or retrolisthesis at the involved segments
- Osteophyte formation on posterior vertebral endplates producing symptomatic
- nerve root or Spinal cord compression and Segmental instability



FINE GRAINED TITANIUM Integration layer of 250 nm

NAKED PEEK ZONE

For easy introduction of the final impactor

**ORTHOBION'S ENGINEERED SURFACE** 

With supporting improved bone cell anchorage and adhesion to biomaterial surfaces

LATERAL HOLES

For graft vascularization

FGOIC-TI TECHNOLOGY AS THE OUTER LAYER


For optimal cell-response and osseointegration & reduced risk of infection

BULLET SHAPE For easy insertion

#### **PRECAUTION FOR USE:**

Implants and instruments must only be used by qualified professionals.

The placement and fitting of the TSC PLIF / TLIF cage must be systematically accompanied with a posterior fixation system.





#### **LARGE WINDOW** For bone craft placement providing increased fusion area

#### **OPTIMAL GUIDANCE** With channel system

#### ANATOMICAL SHAPE & LORDOSIS Allowing a more natural fit

# POSITIONING OF THE PATIENT






The patient is positioned in a standard ventral decubitus position in cases of posterior lumbar interbody fusion. Avoiding compression points and make sure the abdomen is free positioned to avoid pressure on the large vessels and to minimize blood loss.

The intervention is carried out by a posterior median route, X-ray shall be used to determine the correct level and to confirm identification of the affected disc and in a later stage the correct positioning of the implants.

#### Transforaminal approach:

A midline incision is performed over the level to be instrumented. Expose the interlaminair window and the medial parts of the facet joint at unlateral side. Using a combination of surgical instruments a unilateral facetectomy in proformed to provide acces to the disc space. Due to the design (10 mm width and bullet shap nose ) of the TSC TLIF Banana cage can be carfully inserted into the disk space maintaining an angulation of 35-40°.



#### Preparation of intervertebral disc space:

The dura and upper nerve root are carefully protected using nerve root retractors (REF 99.122).

IMPORTANT: Orthobion Instruments are not designed or indicated to be in direct contact with the central nervous system or central circulation system. To retract the central nervous system please use Instruments with this indication!



The quick connect T-Handle (REF 99.005) is required for using the distractors/disc shavers (99.012-99.018) The distraction of the disc space is carried out in turning the distractor in contra clockwise direction.





#### Cleaning the intervertebral disc space:

This can be performed with the available curettes



Rectangular curette REF 99.123



Spoon curette REF 99.124



Ring curette REF 99.125




Down push curette REF 99.120

#### Preparing the endplates:

The distractor/disc shavers (REF 99.012-99.018) are used in a clockwise turning direction.









#### Side Note

The clockwise turning direction results in a sharp disc shaver function, (cleaning and shaving of the endplates) The contra clockwise turning direction of the distractors results in a blunt Distraction. (To create a final disc height estimate for the implant trial height)Those distractors have a double function in function of their turning direction !

(The sizing to determine the final height and length of the TSC PLIF / TLIF cage is been done with the trial cage holder REF 99.311) in combination with the trial cages (REF 99.151-99.164)

Trial cages are available in standard Lordosis 6 ° all heights and lengths 28 mm in all available heights



#### Final TSC Cage impaction:

\* Explanation of the final Cage Impactor




Final TLIF Banana cage inserter 99.006

Introduce Part C (REF 99.006) into Part A then screw the top of the threat inserter to assembly both instruments to one piece. Pull back Part C and connect it to the Cage. Then release Part C and the Cage is locked. If you want to change the angle of the Cage Pull back Part C turn the Inserter.



Cage insertion The TSC PLIF cages have 3 radiopaque titanium markers, here shown as red dots (posterior and the anterior side of the cage). The TSC TLIF has 3 radiopaque tantalum markers. This allows verifying and visualization of the final position of the TSC Cage with per operative image intensifier.









Gottlieb - Dalmler - Strasse 5 D - 78467 Konstanz T. +49 (0) 7531 12218 30 F. +49 (0) 7531 12218 59 E. andreas.ottenschlaeger@orthobion.com

#### www.orthobion.com



ATION BW INNOV 2013

INNOVATION AWARD BADEN-WURTTEMBERG DR.-RUDOLF-EBERLE-PREIS

#### SPECIAL AWARD WINNER 2013

**CE** 0044